



## SWIMMING POOL LIFT

DISPOSITIVO

# i-swim3

MODELLO



**ORIGINAL MANUAL** 



REVISIONE

08/2021

ANNO





# **TABLE OF CONTENT**

Pool Lift

DISPOSITIVO

i-swim3

MODELLO

1.	GENERAL WARNINGS	3
1.1	WARNING	3
1.2	IDENTIFICTION AND POSITIONING	3
2.	GENERAL DESCRIPTION	4
2.1	SOCKET POSITIONING AND INSTALLATION	4
2.2	INSTALLATION AND REMOVAL OF THE LIFT	5



### **GENERAL WARININGS**

**Pool Lift** DISPOSITIVO i-swim3 MODELLO

#### WARNING



THE SOCKET IS A SAFETY COMPONENT AND HAS TO BE **INSTALLED BY SKILLED PERSONS ONLY** 



AFTER INSTALLATION THE LIFT HAS TO BE TESTED IN ORDER TO VERIFY THAT THE SOCKET IS ABLE TO HOLD SAFELY THE MAX LOAD

#### **IDENTIFICATION AND POSITIONING**

The minimum distance of the lift to the pool edge or wall is shown in figure "A". Please position the lift at this suggested distance. The model "standard" and "Maxi" have to be positioned at 220mm and 250mm




Fig. A .ັ ພ<sub>ອ</sub>. J di **5** 



### **GENERAL DESCRIPTION**

Pod	ol Lift
DISF	POSITIVO
i-sv	vim3
MOL	DELLO

#### 2.1 SOCKET INSTALLATION

Figure "B" shows the hole and the positioning details:

- 1. Drill the hole with a hole saw bit Ø100mm for a depth of 170mm.
- 2. Remove the material and prepare the envelope to fit the socket
- 3. Install the socket using concrete Class 32.5 (or higher)
- 4. Position the socket at the centre of the hole, making sure that its upper edge is at the level of the floor , where it is installed. The plastic cover has to protrude by 2mm from the ground (see detail "C")




Fig. B



## **GENERAL DESCRIPTION**

Pool Lift	
DISPOSITIVO	
i-swim3	
MODELLO	

## 2.2 LIFT INSTALLATION AND REMOVAL

As soon as the concrete is solid, remove the socket cover and the lift can be installed

When the lift is removed from the socket close the hole using the plastic cover. The pulling handle can be remove to avoid that somebody trips over it.

